



WASHINGTON STATE PATROL  
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: JAN. 4 2011 INITIALS: D/L OWNER: CITY OF TACOMA

DISTRICT: ADDRESS:

LOCATION: COUNTY:

MAKE: HAENNI TYPE PORT. WHEEL SCALE ACCURACY CLASS: IIII

MODEL: 101 ID. NO: 21964 CAPACITY: 20000 lbs. SECTIONS:

MINIMUM GRADUATIONS: 50 lbs. REG. ELEMENT: DIAL PLATFORM 15"X26"

| LOAD POSITION | TEST WEIGHT | SCALE INDICATION | ERROR | LOAD POSITION | TEST WEIGHT | SCALE INDICATION | ERROR |
|---------------|-------------|------------------|-------|---------------|-------------|------------------|-------|
| CENTER        | 5000        | 5000             | 0     |               |             |                  |       |
| CENTER        | 10000       | 10000            | 0     |               |             |                  |       |
| CENTER        | 15000       | 15000            | 0     |               |             |                  |       |
| CENTER        | 20000       | 20000            | 0     |               |             |                  |       |

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report and Certificate of Calibration", Certification number(s): 36293-W-02

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: Dan J. LaFreniere

Date: Jan. 20, 2011

Print name: Dan LaFreniere

State of Washington

County of Shelton

Signed and sworn to (or affirmed before me on Jan 20, 2011 by

Brenda A Biscay

(Signature of notary public)  
Brenda A. Biscay

Printed Name of Notary

My appointment expires 3.12.12

